

All interpretations are opinions based on inferences from electrical or other measurements and we cannot and do not guarantee the accuracy or correctness of any interpretation and we shall not, except in the case of gross or willful negligence on our part, be liable or responsible for any loss, costs, damages or expenses incurred or sustained by anyone resulting from any interpretation made by any of our officers, agents or employees. These interpretations are also subject to our General Terms and Conditions as set out in our current Price Schedule.

Comments

A-1 CALIPER
GAMMA RAY



Dual Induction Calibration Report

Serial Model:
Date Last Performed: Sat Jan 12 11:15:16 2007
Checked / as Performed: Sat Jan 12 12:56:39 2008
Next Survey Due/Status Performed: Sat Jan 12 12:56:39 2008

Surface Calibration

	Readings			References		Results	
Loop	Air	Loop		Air	Loop	Air	h
Deep	41.656	0.047	V	0.000	400.000 mmho-m	812.500	0.301
Medium	41.656	0.774	V	0.000	400.000 mmho-m	805.356	4.261
Internal	Zero	Cal		Zero	Cal	m	b
Deep	0.041	0.047	V	0.000	400.000 mmho-m	812.500	0.301
Medium	0.000	0.733	V	0.000	387.500 mmho-m	805.356	4.274

Downhole Calibration

	Readings			References		Results	
Internal	Zero	Cal		Zero	Cal	m	b
Deep	0.000	400.102	mmho-m	0.726	399.000 mmho-m	0.981	-0.709
Medium	0.000	475.316	mmho-m	1.000	481.900 mmho-m	0.972	-1.104
Shallow	0.000	0.000	mmho-m	0.000	68.400 mmho-m	0.000	0.000

After Survey Calibration

	Readings			Targets		Results	
Internal	Zero	Cal		Zero	Cal	mf	bf
Deep	10.208	400.014	mmho-m	9.994	409.700 mmho-m	0.981	-9.709
Medium	0.000	475.316	mmho-m	1.000	481.900 mmho-m	0.972	-1.104
Shallow	0.000	0.000	mmho-m	0.000	68.400 mmho-m	1.027	-0.036